


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>0100</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<p><b>Johnson Matthey plc - Noble Metals</b></p> <p>Issue No: 016    Issue date: 21 July 2010</p>	
	<p>Orchard Road Royston Herts SG8 5HE</p>	<p>Contact: Nick Mahoney Tel: +44(0) 1763 253893 Fax: +44 (0)1763 253451 E-Mail: <a href="mailto:nick.mahoney@matthey.com">nick.mahoney@matthey.com</a> Website: <a href="http://www.noble.matthey.com">www.noble.matthey.com</a></p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
TEMPERATURE			
Platinum thermocouples			
Fixed point measurements	0 °C to 1100 °C 1100 °C to 1600 °C 1600 °C to 1800 °C	1.0 °C 2.0 °C 3.0 °C	
Measurement by comparison	0 °C to 1100 °C 1100 °C to 1400 °C	2.0 °C 3.0 °C	
<p>END</p>			