


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0115 Accredited to ISO/IEC 17025:2005	Lion Laboratories Limited Issue No: 013 Issue date: 23 July 2010	
	Calibration Laboratory Ty Verlon Industrial Estate Barry South Glamorgan Wales CF63 2BE	Contact: Mr R D Blyth Tel: +44 (0)1446-724500 Fax: +44 (0)1446-724501 E-Mail: dblyth@lionlaboratories.com Website: www.lionlaboratories.com

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Calibration Laboratory Ty Verlon Industrial Estate Barry South Glamorgan Wales CF63 2BE	Local contact Mr D R Blyth	Evidential Breath Testing Instruments	A

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Customers Sites	Contact: Mr D R Blyth	Evidential Breath Testing Instruments	B



0115
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Lion Laboratories Limited
Issue No: 013 Issue date: 23 July 2010

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration & Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
CALIBRATION OF EVIDENTIAL BREATH TESTING INSTRUMENTS Lion Intoxilyzer 6000	0 µg to 200 µg ethanol/100 ml simulated breath	1.0 µg ethanol/100 ml simulated breath	Evidential breath testing instruments manufactured by Lion Laboratories Ltd approved by the Home Secretary	A
	Effect of interfering substances acetone methanol methyl ethyl ketone toluene		Confirmation of compliance with Home Office Quality Framework Document	A
	0 µg to 140 µg ethanol/100 ml simulated breath	1.0 µg ethanol/100 ml simulated breath		A & B
	Effect of interfering substances Acetone 15 µg/100 ml Methanol 4 µg/100 ml		Confirmation of compliance with Home Office Quality Framework Document	A & B
CALIBRATION OF MOBILE EVIDENTIAL BREATH TESTING INSTRUMENTS Lion Intoxilyzer 8000	0 µg to 200 µg ethanol/100 ml simulated breath	1.0 µg ethanol/100 ml simulated breath	Mobile evidential breath testing instruments manufactured by Lion Laboratories Ltd not currently approved by the Home Secretary	A
	Effect of interfering substances acetone methanol methyl ethyl ketone toluene		Confirmation of compliance with Home Office Quality Framework Document	A
	0 µg to 140 µg ethanol/100 ml simulated breath	1.0 µg ethanol/100 ml simulated breath		A
	Effect of interfering substances Acetone 15 µg/100 ml Methanol 4 µg/100 ml		Confirmation of compliance with Home Office Quality Framework Document	A
END				