

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>0136</p> <p>Accredited to ISO/IEC 17025:2005</p>	BSI Assurance UK Limited Issue No: 015 Issue date: 18 January 2012	
	BSI Kitemark Court Davy Avenue Knowlhill Milton Keynes MK5 8PP	Contact: Ms Natalie Bennett Tel: +44 (0)1908 814707 Email: natalie.bennett@bsigroup.com Sales Enquiry Email: Certification.Sales@bsigroup.com Sales Enquiry Tel: +44 (0)845 080 9000 Fax: +44 (0)1908 228060 Website: www.bsigroup.com
Calibration performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Location details	Activity	Location code
<p>Address BSI Kitemark Court Davy Avenue Knowlhill Milton Keynes MK5 8PP</p> <p>Local contact Ms Natalie Bennett Tel: +44 (0)1908 814707 Fax: +44 (0)1908 228060 Email: natalie.bennett@bsigroup.com Website: www.bsigroup.com</p>	<p>Support Functions: Quality Management including document control, auditing and quality control</p>	A
<p>Address Photometric Measurement and Calibration Department Lighting and Photometry Laboratories Maylands Avenue Hemel Hempstead Hertfordshire HP2 4SQ</p> <p>Local contact Mr Chris Higby Tel: +44 (0)1442 278525 Fax: +44 (0)1442 278531 Email: chris.higby@bsigroup.com Website: www.bsigroup.com</p>	<p>Photometric Calibration: Luminous Flux, Luminous Intensity and Illuminance</p>	B



0136
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

BSI Assurance UK Limited
Issue No: 015 Issue date: 18 January 2012

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
LUMINOUS FLUX Tungsten filament lamps	0.5 to 10 lumen 10 to 20 000 lumen	2.0 % 1.0 %		B B
LUMINOUS INTENSITY Tungsten filament lamps	0.5 to 5 000 candela	1.5 %		B
ILLUMINANCE Illuminance meters Light meters Lux meters Photometers White light meters	2 to 2 000 lux 2 000 to 10 000 lux	2.0 % 2.5 %		B B
END				