


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0148 Accredited to ISO/IEC 17025:2005	T&J Farnell	
	Issue No: 024 Issue date: 06 August 2010	
	Unit 14 Walkers Road North Moons Moat Redditch Worcestershire B98 9HE	Contact: Mr T Rankin Tel: +44 (0)1527 758 737 Fax: E-Mail: sales@tjfarnell.co.uk Website:
Calibration performed at the above address only		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
SURFACE MEASUREMENT TRRL Rolling Straightedge	Road Surface Depressions 0 mm to 12 mm Odometer Wheel Diameter 115 mm to 120 mm	0.20 mm 0.090 mm	
END			