

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;"><b>0179</b></p> <p style="text-align: center;">Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Michell Instruments Ltd</h3> <p><b>Issue No: 019    Issue date: 11 January 2011</b></p>	
	<p><b>48 Lancaster Business Park</b> Ely Cambridgeshire CB6 3NW</p>	<p><b>Contact: Mr C Parker/Mr P Newton</b> <b>Tel: +44 (0)1353 658000</b> <b>Fax: +44 (0)1353 658199</b> <b>E-Mail: info@michell.co.uk</b> <b>Website: www.michell-instruments.com</b></p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
HUMIDITY			
Dewpoint	90 °C to 82 °C 82 °C to 20 °C 20 °C to - 60 °C - 60 °C to - 75 °C - 75 °C to - 90 °C	0.30 °C to 0.24 °C 0.24 °C 0.15 °C to 0.24 °C 0.24 °C to 0.40 °C 0.40 °C to 0.80 °C	Calibrations can be performed on instruments with an electrical output
Relative humidity			
Relative humidity instruments	10 %rh to 98 %rh, 20 °C to 90 °C 10 %rh to 98 %rh, 10 °C to 20 °C	2.3 % of reading 2.6 % of reading	
Temperature sensors incorporated in humidity instruments	10 °C to 90 °C	0.23 °C	
TEMPERATURE			
Electronic thermometers with sensors	0 °C to 90 °C	0.17 °C	
END			