


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;">UKAS CALIBRATION</p> <p style="text-align: center;">0260</p> <p style="text-align: center;">Accredited to ISO/IEC 17025:2005</p>	<h3>Norfolk Calibration Services</h3> <p>Issue No: 022 Issue date: 6 May 2011</p>	
	<p>Norfolk County Council Environment, Transport and Development Trading Standards County Hall Norwich NR1 2DH</p>	<p>Contact: Mr Adrian Chapman Tel: +44 (0)1603-222177 Fax: +44 (0)1603-222472 E-Mail: calibration@norfolk.gov.uk Website: http://www.norfolk.gov.uk/calibration</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Norfolk County Council Environment, Transport and Development Trading Standards County Hall Norwich NR1 2DH</p> <p>Local contact Adrian Chapman</p>	<p>Mass (Weights) Mass (Weighing machines)</p>	<p>Lab</p>

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>Customers' Premises The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p>Mass (Weighing machines)</p>	<p>Site</p>



0260
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Norfolk Calibration Services
Issue No: 022 Issue date: 6 May 2011

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
MASS See Notes 1, 2 and 3	Nominal value(g)	(mg)	<p>1 Intermediate values can be calibrated with an uncertainty equal to that of the next higher nominal value.</p> <p>2 Calibrations can be given in other units as required.</p> <p>3 The Best Measured Capability shown will permit statements of compliance to the Maximum Permissible Errors shown in OIML R111 up to Class F1 at 25kg, and up to Class E2 from 20kg to 1mg.</p>	Lab
	25 000	21		
	20 000	7.0		
	10 000	4.2		
	5 000	2.0		
	2 000	0.70		
	1 000	0.40		
	500	0.18		
	200	0.080		
	100	0.040		
	50	0.028		
	20	0.015		
	10	0.012		
	5	0.008 0		
	2	0.007 0		
	1	0.007 0		
	0.5	0.004 5		
	0.2	0.004 5		
	0.1	0.003 8		
	0.05	0.003 6		
	0.02	0.002 4		
0.01	0.001 6			
0.005	0.001 6			
0.002	0.001 6			
0.001	0.001 6			
NON AUTOMATIC WEIGHING MACHINES See notes 4 and 5	1 g	0.015 mg	<p>4. Weights are available in OIML Class:</p> <p>E2 from 1 mg to 1 kg, max. grouped load 2.5 kg</p> <p>F1 from 1 mg to 20 kg, max. grouped load 75 kg</p> <p>M1 from 500 mg to 20 kg, max. grouped load 2300 kg</p> <p>5. Other loads within the overall listed range may also be used.</p>	Lab & Site
	2 g	0.018 mg		
	5 g	0.023 mg		
	10 g	0.032 mg		
	20 g	0.044 mg		
	50 g	0.095 mg		
	100 g	0.13 mg		
	200 g	0.27 mg		
	500 g	0.69 mg		
	1 kg	1.3 mg		
	2 kg	2.7 mg		
	2.5 kg	3.4 mg		
	5 kg	8.4 mg		
	10 kg	19 mg		
	20 kg	38 mg		
	50 kg	110 mg		
	75 kg	150 mg		
	100 kg	840 mg		
	200 kg	3.7 g		
	500 kg	9.2 g		
1000 kg	19 g			
2000 kg	43 g			
2300 kg	70 g			