


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>Accredited to ISO/IEC 17025:2005</p>	<h3>West Yorkshire Calibration Laboratory</h3> <p>Issue No: 021 Issue date: 05 September 2011</p>	
	<p>Nepshaw Lane South Morley Leeds LS27 0QP</p>	<p>Contact: Mr S Kaye Tel: +44 (0)113 393 9891 Fax: +44 (0)113 253 0311 E-Mail: metrology@wyjs.org.uk Website: http://www.calibration.wyjs.org.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
MASS	kg	g	NOTES 1 Intermediate values can be calibrated to an uncertainty proportional between those of the next higher and lower nominal values. 2 Calibrations can be given in other units by arrangement.
Nominal values See Note 1	1000 500 250 200 100 50	20.0 10.0 5.0 4.0 2.0 2.0	
Nominal values See Note 1	g	mg	
	25 000 20 000 10 000 5 000 2 000 1 000 500 200 100 50 20 10 5 2 1 0.5 0.2 0.1 0.05 0.02 0.01 0.005 0.002 0.001	12.5 10.0 5.0 2.5 1.0 0.50 0.25 0.10 0.050 0.033 0.027 0.020 0.017 0.013 0.010 0.0080 0.0070 0.0050 0.0040 0.0030 0.0030 0.0020 0.0020 0.0020	
END			