

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>0308</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Fife Council</h3> <p>Issue No: 012 Issue date: 04 August 2010</p>	
	<p>Environmental Services Trading Standards Kingdom House, Kingdom Avenue Glenrothes Fife Scotland KY7 5LY</p>	<p>Contact: Mr R A Hamilton Tel: +44 (0)8451 555555 ext 440332 Fax: +44 (0)1592 583318 E-Mail: alan.hamilton@fife.gov.uk Website:</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
MASS	Nominal value (g)	(mg)	Intermediate values can be calibrated with an uncertainty not less than that interpolated from the next higher and lower nominal value in the table
	25 000	300	
	20 000	60	
	10 000	30	
	5 000	15	
	2 000	6.0	
	1 000	3.0	
	500	1.5	
	200	0.60	
	100	0.30	
	50	0.20	
	20	0.16	
	10	0.12	
	5	0.10	
	2	0.080	
	1	0.060	
	0.5	0.050	
	0.2	0.040	
	0.1	0.030	
	0.05	0.024	
	0.02	0.020	
	0.01	0.016	
	0.005	0.012	
	0.002	0.012	
	0.001	0.012	
END			