

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION</p> <p>0313</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p><b>Linear Instrument Services Ltd</b></p> <p>Issue No: 019    Issue date: 10 June 2010</p>	
	<p>6A St Francis Way Shefford Bedfordshire SG17 5DZ</p>	<p>Contact: Mr J J Wiltshire Tel: +44 (0)1462-851177 Fax: +44 (0)1462-851180 E-Mail: johnjwiltshire@aol.com Website:</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> 6A St Francis Way Shefford Bedfordshire SG17 5DZ</p> <p><b>Local contact</b> Mr J Wiltshire</p>	<p><b>Calibration:</b>  Electrical calibration of temperature simulators, indicators, controllers and recorders</p>	<p>Shefford</p>

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p><b>Customers' sites or premises</b></p> <p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p><b>Calibration:</b>  Electrical calibration of temperature simulators, indicators, controllers and recorders</p>	<p>Customers' Sites</p>



0313  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Linear Instrument Services Ltd**  
**Issue No: 019    Issue date: 10 June 2010**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Electrical calibration of temperature indicators, controllers and recorders for the following sensors:				Sheffield and Customers' Sites
Noble metal thermocouples	- 50 °C to + 1750 °C	1.4 °C	Including cold junction compensation	
Base metal thermocouples	- 200 °C to + 1370 °C	1.1 °C	Including cold junction compensation	
END				