

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

| | | |
|---|--|---|
|  <p>UKAS CALIBRATION 0315</p> <p>Accredited to ISO/IEC 17025:2005</p> | <p>Warwickshire County Council</p> <p>Issue No: 014 Issue date: 10 January 2012</p> | |
| | <p>Trading Standards Service Old Budbrooke Road Warwick CV35 7DP</p> | <p>Contact: Mr C Hayter Tel: +44 (0)1926 414076 Fax: +44 (0)1926 414014 E-Mail: clivehayter@warwickshire.gov.uk Website: www.warwickshire.gov.uk</p> |
| <p>Calibration performed at the above address only</p> | | |

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks |
|--|---|---|--|
| <p>FLOW</p> <p>Hydrocarbon oils Quantity of fluid passed</p> | <p>1000 & 5000 litres at flow rates from 35 l min⁻¹ to 5000 l min⁻¹</p> | <p>0.050 % of measured quantity</p> | <p>1 Calibrations can be carried out using SBP11, Kerosene, or gas-oil.</p> <p>2 Calibrations are carried out at ambient temperature and converted to 15 °C.</p> <p>3 Calibrations are carried out at pressures of up to 2 bar.</p> <p>4 All calibrations can be given in metric, imperial or other units as required.</p> |
| <p>END</p> | | | |