


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>0320</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Suffolk County Council</h3> <p>Issue No:015 Issue date: 30 July 2010</p>	
	<p>Trading Standards Laboratory St Edmund House Rope Walk Ipswich Suffolk IP4 1LZ</p>	<p>Contact: Mr A J Runnacles Tel: +44 (0)1473-584288 E-Mail: john.runnacles@suffolk.gov.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
MASS	Nominal value (g)	(mg)	NOTES: 1 Intermediate values can be calibrated to an uncertainty equal to that of the next higher nominal value. 2 Calibrations can be given in other units as required.
	25 000	25	
	20 000	20	
	10 000	10	
	5 000	5.0	
	2 000	2.0	
	1 000	1.0	
	500	0.50	
	200	0.20	
	100	0.10	
	50	0.060	
	20	0.050	
	10	0.040	
	5	0.030	
	2	0.020	
	1	0.020	
	0.5	0.016	
	0.2	0.012	
	0.10	0.010	
	0.05	0.0080	
	0.02	0.0060	
	0.01	0.0050	
	0.005	0.0040	
	0.002	0.0040	
	0.001	0.0040	
END			