


Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0327 Accredited to ISO/IEC 17025:2005	Bowmonk Limited Issue No: 011 Issue date: 05 April 2006	
	Diamond Road St Faiths Industrial Estate Norwich Norfolk NR6 6AW	Contact: Mr E Hill Tel: +44 (0)1603-485153 Fax: +44 (0)1603-418150 E-Mail: info@bowmonk.co.uk Website: www.bowmonk.com
Calibration performed at the above address only		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks
Portable Accelerometer (brake meters)	Horizontal acceleration expressed as 0 to 100% of the local acceleration due to gravity	0.2% at 45 degrees and over range of calibration	NOTES 1 All calibrations must be carried out in accordance with procedures agreed by UKAS. 2 Uncertainties quoted are associated with the setting of the calibration equipment and do not take into account the characteristics of the device being calibrated.
END			