


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0327 Accredited to ISO/IEC 17025:2005	Bowmonk Limited	
	Issue No: 012 Issue date: 4 August 2010	
	Diamond Road St Faiths Industrial Estate Norwich Norfolk NR6 6AW	Contact: Mr E Hill Tel: +44 (0)1603-485153 Fax: +44 (0)1603-418150 E-Mail: info@bowmonk.co.uk Website: www.bowmonk.com
Calibration performed at the above address only		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
Portable Accelerometer (brake meters)	Horizontal acceleration 5 % to 100%	0.20 %	NOTE Horizontal acceleration expressed as a percentage (%) of the local free fall acceleration due to gravity.
END			