

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>0349</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Cardiff Metrology Services</h3> <p>Issue No:011 Issue date: 11 August 2010</p>	
	<p>Environmental Protection</p> <p>County Hall Atlantic Wharf Cardiff CF10 4UW</p>	<p>Contact: Jacqueline James</p> <p>Tel: +44 (0)29-20-872069 Fax: +44 (0)29-20-872072 E-Mail: jjames@cardiff.gov.uk Website:</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
MASS	(g)	(mg)	<ol style="list-style-type: none"> Intermediate values can be calibrated to an uncertainty interpolated from the next higher and lower values in the table. Calibrations can be given in other units as required.
	25 000	75	
	20 000	60	
	10 000	30	
	5 000	15	
	2 000	6.0	
	1 000	3.0	
	500	1.5	
	200	0.60	
	100	0.30	
	50	0.20	
	20	0.16	
	10	0.12	
	5	0.10	
	2	0.080	
	1	0.060	
	0.5	0.050	
	0.2	0.040	
	0.1	0.030	
	0.05	0.024	
	0.02	0.020	
	0.01	0.016	
	0.005	0.012	
	0.002	0.012	
	0.001	0.012	
END			