

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION</p> <p>0352</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Kent Scientific Services</h3> <p>Issue No: 024    Issue date: 21 December 2010</p>	
	<p>8 Abbey Wood Road Kings Hill West Malling Kent ME19 4YT</p>	<p>Contact: Mr K Hutchins Tel: +44 (0)1732-220001 Fax: +44 (0)1732-220006 E-Mail: kim.hutchins@kent.gov.uk Website: www.kss-labs.co.uk</p>
<p>Calibration performed by the Organisation at the locations specified below</p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> 8 Abbey Wood Road Kings Hill West Malling Kent ME19 4YT</p> <p><b>Local contact</b> Mr K Hutchins</p>	<p>Mass (weights) Temperature</p>	<p>Lab</p>

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>Customers' Premises.</p> <p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p>Mass (Weighing machines)</p>	<p>Site</p>



0352  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Kent Scientific Services**  
**Issue No: 024 Issue date: 21 December 2010**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks	Location Code
MASS See Notes 1 and 2	Nominal value (g)	(mg)	NOTES	Lab
Specific values	30 000 25 000 20 000 10 000 5 000 2 000 1 000 500 200 100 50 20 10 5 2 1 0.5 0.2 0.1 0.05 0.02 0.01 0.005 0.00 0.001	300 250 20 10 5.0 2.0 1.0 0.50 0.20 0.10 0.060 0.050 0.040 0.030 0.024 0.020 0.016 0.012 0.010 0.008 0 0.006 0 0.005 0 0.004 0 0.004 0 0.004 0	1 Calibrations can be given in other units as required.  2 Intermediate values of weights can be calibrated with an uncertainty not less than that interpolated from the next higher and next lower nominal values.	
Non-specific values	From 50 kg to 1300 kg	20 ppm		



0352  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Kent Scientific Services**  
**Issue No: 024 Issue date: 21 December 2010**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks	Location Code
NON AUTOMATIC WEIGHING MACHINES  Electronic, single pan (See Notes 1 and 2)	10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg  5 kg 10 kg 20 kg 40 kg  50 kg 100 kg 200 kg 500 kg 1000 kg 1150 kg	0.033 mg 0.046 mg 0.77 mg 0.14 mg 0.28 mg 0.69 mg 1.4 mg 2.8 mg  9.6 mg 20 mg 39 mg 77 mg  0.77 g 1.7 g 4.3 g 9.7 g 20 g 22 g	1. Weights are available in OIML Class:  E2 from 1 g to 1 kg, max. grouped load 2 kg  F1 from 1 g to 20 kg, max. grouped load 40 kg  M1 from 1 g to 20 kg, max. grouped load 1150 kg  2. Other loads within the overall listed range may also be used.	Lab & Site
TEMPERATURE  Liquid-in-glass thermometers	-70 °C to -40 °C -40 °C to 100 °C 100 °C to 250 °C	0.070°C 0.050°C 0.070°C		Lab
Electronic thermometer with indicators	-70 °C to -40 °C -40 °C to 100 °C 100 °C to 250 °C	0.070°C 0.050°C 0.070°C		Lab

END