

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="margin: 0;"><b>0405</b></p> <p style="margin: 0;">Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3 style="margin: 0;">Gemco Service Limited</h3> <p style="margin: 0;">Issue No: 026    Issue date: 18 October 2010</p>	
	<p>Road One Winsford Industrial Estate Winsford Cheshire CW7 3PL</p>	<p>Contact: Mr N Blake Tel: +44 (0)1606 542321 Fax: +44 (0)1606 861853 E-Mail: <a href="mailto:nick.blake@gemco.co.uk">nick.blake@gemco.co.uk</a> Website: <a href="http://www.gemco.co.uk">www.gemco.co.uk</a></p>
<p><b>Calibration performed by the Organisations at the locations specified below</b></p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory Locations:

Location details		Activity	Location code
<p><b>Address</b> Road One Winsford Industrial Estate Winsford Cheshire CW7 3PL</p>	<p><b>Local contact</b> Mr N Blake  Tel: +44 (0)1606 542321 Fax: +44 (0)1606 861853 Email: <a href="mailto:nick.blake@gemco.co.uk">nick.blake@gemco.co.uk</a> Website: <a href="http://www.gemco.co.uk">www.gemco.co.uk</a></p>	<p>Temperature (Electronic thermometers with sensors) Pressure</p>	<p>P  P</p>

#### Site activities performed away from the locations listed above\*:

Location details	Activity	Location code
<p>Customer Premises eg, MoT Test Centres, Garages, and Repair/Service Facilities (*including those of Gemco Service, Winsford)</p>	<p>Vehicle Exhaust Analyser (VEGA/DSM) Calibrations</p>	<p>S</p>



0405  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Gemco Service Limited**  
**Issue No: 026 Issue date: 18 October 2010**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
Instruments approved by the Vehicle Inspectorate on behalf of the DfT for the measurement of:-	As per Vehicle Inspectorate requirements:-		A list of individual approved signatories and the type of approved instruments they may calibrate is held by the laboratory and by UKAS.	
a) vehicle exhaust emissions	VPB/07/24/20/CAL dated May 1995 (1 <sup>st</sup> revision August 1995)	3.0 % of reading		S
b) free acceleration smoke	MOT/08/19/01 dated October 2001 (3 <sup>rd</sup> revision January 2007)	0.10 m <sup>-1</sup>	Smoke obscuration coefficient	S
<b>TEMPERATURE</b>				
Electronic thermometers with sensors, and temperature indicators	- 10 °C to + 100 °C	0.15 °C		P
<b>PRESSURE</b>				
Gas pressure (absolute)				
Calibration of indicating instruments and gauges	12 kPa to 202 kPa	130 Pa		P
Gas pressure (gauge)				
Calibration of indicating instruments and gauges	- 88 kPa to 102 kPa	75 Pa		P
END				