


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION</p> <p><b>0420</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<b>CERAM</b>	
	<b>Issue No: 015      Issue date: 3 August 2011</b>	
	<b>CERAM Research Ltd</b> Queens Road Penkhull Stoke-on-Trent ST4 7LQ	<b>Contact: Lee Bullock</b> <b>Tel: +44 (0)1782-764444</b> <b>Fax: +44 (0)1782-412331</b> <b>E-Mail: lee.bullock @ceram.com</b> <b>Website: www.ceram.com</b>
<b>Calibration performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( <i>k</i> =2)	Remarks
Spectral Reflectance Curves and derived Colorimetric Data. Direct Comparison with NPL Standard.	Reflectance: 0 – 100 % Wavelength range: 380 nm to 750 nm	Reflectance High values 0.81 % to Low values 0.21 %	
Spectral Reflectance Curves and derived Colorimetric Data. Absolute Measurement.	Reflectance: 0 – 100 % Wavelength range: 380 nm to 750 nm	Reflectance High values 0.97 % to Low values 0.28 %	
END			