

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>0430</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<b>D Brash &amp; Sons Ltd</b>	
	<b>Issue No: 017</b>	<b>Issue date: 6 December 2010</b>
<b>37 Stamperland Crescent</b> <b>Clarkston</b> <b>Glasgow</b> <b>G76 8LH</b>	<b>Contact: Mr D G Christie</b> <b>Tel: +44 (0)141-638 2284</b> <b>Fax: +44 (0)141-620 1842</b> <b>E-Mail: service@dbrash.co.uk</b> <b>Website:</b>	
<b>Calibration performed by the Organisations at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	
<b>Address</b> 37 Stamperland Crescent Clarkston Glasgow G76 8LH	<b>Local contact</b> Mr D G Christie Tel: +44 (0)141-638 2284 Fax: +44 (0)141-620 1842 E-Mail: service@dbrash.co.uk
<b>Address</b> Unit 7, Slough Business Centre Bristol Way Slough Berkshire SL1 3TD	<b>Local contact</b> Mr G Emerton Contact: Mr G Emerton Tel: +44 (0) 1753 511 801 Fax: +44 (0) 1753 694 447 E-Mail: Sales@scaleservices.co.uk

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers' sites or premises  The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	Mass – weighing machines (non-automatic)	S



0430  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**D Brash & Sons Ltd**  
**Issue No: 017 Issue date: 06 December 2010**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks	Location Code
NON AUTOMATIC WEIGHING MACHINES	200 mg 500 mg  1g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g  1 kg 2 kg 5 kg 10 kg 20 kg 50 kg 60 kg 100 kg 200 kg 500 kg 1000 kg 1500 kg 2000 kg 4500 kg	0.015 mg 0.020 mg  0.025 mg 0.030 mg 0.038 mg 0.051 mg 0.067 mg 0.095 mg 0.17 mg 0.34 mg 0.85 mg  1.7 mg 3.4 mg 9.5 mg 19 mg 39 mg 110 mg 130 mg 1.7 g 4.4 g 10 g 20 g 26 g 44 g 90 g	1. Weights are available in OIML Class  E2 from 1 mg to 500 g Max grouped load 2 kg  F1 from 1 g to 20 kg Max grouped load 60 kg  M1 from 1 kg to 60 kg Max. grouped load 1500 kg  2. Make-up weights may be used for calibration above 300 kg, and will be used for calibrations above 1500 kg up to a maximum of 4500 kg.  3. Other loads within the overall listed range may also be used	S
END				