

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>0440</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3 style="margin: 0;">Wortley Standards Limited</h3> <p style="margin: 0;"><b>Issue No: 013    Issue date: 28 November 2011</b></p>	
	<p><b>Unit 10 Compass Court</b>  <b>Westthorpe Fields Business Park</b>  <b>Killamarsh</b>  <b>Sheffield</b>  <b>S21 1TW</b></p>	<p><b>Contact: Mr P Johnson</b>  <b>Tel: +44 (0)114 248 2443</b>  <b>Fax: +44 (0)114 247 3153</b>  <b>E-Mail: info@wortleystandards.co.uk</b>  <b>Website: www.wortleystandards.co.uk</b></p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
<p><b>RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED</b></p>			
<b>LENGTH</b>			<b>NOTES</b>
Plain plug gauges (parallel), cylindrical setting standards, gear measuring cylinders and rollers	1 to 100 diameter 100 to 150 150 to 200 200 to 300	1.0 1.5 2.0 3.0	
Plain ring gauges (parallel) and setting standards	5 to 25 diameter 25 to 50 50 to 100 100 to 150r 150 to 200 200 to 300r	1.0 1.5 2.0 2.5 3.0 3.5	
Plain gap gauges (parallel)	5 to 100 100 to 200 200 to 300	3.0 5.0 8.0	
<b>MEASURING INSTRUMENTS AND MACHINES</b>			
Micrometers External Internal Depth	As BS 870:2008 0 to 1000 As BS 959:2008 0 to 900 As BS 6468:2008 0 to 450	<div style="border-left: 1px solid black; border-right: 1px solid black; padding: 0 10px;">                     Heads: 2.0 between any two points                       Setting and extension rods: 1.0 + (8.0 x length in m)                 </div>	
Micrometer heads	As BS 1734:1951 0 to 100		1.0
Bench micrometer	As NPL MOY/SCMI 22 0 to 100	Overall performance 2.0 between any two points	



0440  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Wortley Standards Limited**  
**Issue No: 013 Issue date: 28 November 2011**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED			
Vernier caliper, height and depth gauges	As BS 887:2008 0 to 1500 As BS 1643:2008 0 to 1200 As BS 6365: 2008 0 to 600	Overall performance $10 + (30 \times \text{length in m})$	
Dial gauges and dial test indicators	As BS 907:2008 and BS 2795:1981 0 to 50	1.0	
END			