


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>0450</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Avery Weigh-Tronix Calibration Service</b>	
	Issue No: 018    Issue date: 15 December 2011	
	Foundry Lane Smethwick Warley West Midlands B66 2LP	Contact: Mr Ishverlal Patel Tel: +44 (0)121 568 1576 Fax: +44 (0)121 697 5576 E-Mail: ipatel@awtxglobal.com Website: www.averyweigh-tronix.com

**Calibration performed by the Organisations at the locations specified below**

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Foundry Lane Smethwick Warley West Midlands B66 2LP  <b>Local contact</b> Mr Ishverlal Patel  Tel: +44 (0)121 568 1576 Fax: +44 (0)121 697 5576	Mass Mass (weights)	Perm Lab

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customers premises  The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Mass (Weighing machines)	Site



0450  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Avery Weigh-Tronix Calibration Service**  
**Issue No: 018    Issue date: 15 December 2011**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
MASS	Nominal value (g)	(mg)		Perm Lab
	25 000	200		
	20 000	20		
	10 000	10		
	5 000	5.0		
	4 500	50		
	2 500	50		
	2 000	2.0		
	1 000	1.0		
	500	0.5		
	200	0.20		
	100	0.10		
	50	0.060		
	20	0.050		
	10	0.040		
	7.5	0.40		
	5	0.032		
	2	0.024		
	1	0.020		
	0.5	0.016		
	0.2	0.012		
	0.1	0.010		
	0.05	0.008 0		
	0.02	0.006 0		
	0.01	0.005 0		
	0.005	0.004 0		
	0.002	0.004 0		
	0.001	0.004 0		



0450  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Avery Weigh-Tronix Calibration Service**  
**Issue No: 018 Issue date: 15 December 2011**

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
NON AUTOMATIC WEIGHING MACHINES See notes 1 and 2	200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg  2 kg 5 kg 10 kg 20 kg 50 kg 100 kg 200 kg  500 kg 1000 kg 2000 kg 5000 kg	0.015 mg 0.020 mg 0.025 mg 0.030 mg 0.038 mg 0.051 mg 0.067 mg 0.095 mg 0.17 mg 0.34 mg 0.89 mg 1.7 mg  5.5 mg 13 mg 27 mg 56 mg 150 mg 700 mg 3.1 g  30 g 60 g 120 g 290 g	1. Weights are available in OIML Class:  E2 from 1 mg to 200 g, max. grouped load 1 kg       F1 from 5 mg to 10 kg, max. grouped load 200 kg       M1 from 1 g to 1000 kg, max. grouped load 5000 kg  2. Other loads within the overall listed range may also be used.	Site
END				