

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK



0458

Accredited to
ISO/IEC 17025:2005

Sandwell Metropolitan Borough Council

Issue No: 011

Issue date: 20 February 2012

Environmental Health and Trading
Standards
Coneygree Road
Tipton
West Midlands
DY4 8XR

Contact: Ms J Moore
Tel:
Fax: +44 (0)121-569 6647
E-Mail: june_moore@sandwell.gov.uk
Website: metrology.sandwell.gov.uk

Calibration performed at the above address only

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
MASS	Nominal value (g)	(mg)	NOTE
Weights and artefacts	25 000	80	1. Calibrations can be given in other units as required.
	20 000	20	
	10 000	10	
	5 000	5.0	2. Intermediate values can be calibrated to an uncertainty interpolated from the next higher and lower values in the table.
	2 000	2.0	
	1 000	1.0	
	500	0.50	
	200	0.20	
	100	0.10	
	50	0.060	
	20	0.050	
	10	0.040	
	5	0.030	
	2	0.024	
	1	0.020	
	0.5	0.016	
	0.2	0.012	
	0.1	0.010	
	0.05	0.0080	
	0.02	0.0060	
	0.01	0.0050	
	0.005	0.0040	
	0.002	0.0040	
	0.001	0.0040	
END			