


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION <b>0459</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Precision Balance Services</h3> <p><b>Issue No:</b> 011    <b>Issue date:</b> 10 November 2011</p>	
	<p><b>3 Atlas Court</b> Atlas Road Hermitage Industrial Estate Coalville Leicestershire LE67 3FL</p>	<p><b>Contact:</b> Mrs L I Exell <b>Tel:</b> +44 (0)1530-834650 <b>Fax:</b> +44 (0)1530-834650 <b>E-Mail:</b> <a href="mailto:service@precisionbalance.co.uk">service@precisionbalance.co.uk</a> <b>Website:</b> <a href="http://www.precisionbalance.co.uk">www.precisionbalance.co.uk</a></p>
<p><b>Calibration performed by the Organisations at the locations specified below</b></p>		

### Locations covered by the organisation and their relevant activities

<u>Location details</u>	<u>Activity</u>	<u>Location code</u>
<p><b>Address</b> 3 Atlas Court Atlas Road Hermitage Industrial Estate Coalville Leicestershire LE67 3FL</p> <p><b>Local contact</b> Mrs L Exell</p>	<p>Portable weighing machines (non-automatic)</p>	<p>P</p>

### Site activities performed away from the locations listed above:

<b>Location details</b>	<b>Activity</b>	<b>Location code</b>
<p>Customer Sites                      Contact: Mrs L I Exell</p> <p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p>	<p>MASS Weighing machines (non-automatic)</p>	<p>S</p>



0459  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Precision Balance Services**  
**Issue No: 011 Issue date: 10 November 2011**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks	Location Code
NON-AUTOMATIC WEIGHING MACHINES See notes 1 and 2.	200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg  10 kg 20 kg 50 kg 100 kg 200 kg 500 kg	0.0092 mg 0.012 mg 0.015 mg 0.019 mg 0.024 mg 0.033 mg 0.045 mg 0.074 mg 0.014 mg 0.28 mg 0.69 mg 1.39 mg 2.77 mg 6.93 mg  13.9 mg 39 mg 110 mg 1.0 g 4.4 g 9.9 g	1. Weights are available in OIML Class:  E2 from 1 mg to 5 kg. Max grouped load 11 kg.  F1 1 mg to 10 kg. Max grouped load 70 kg  M1 5 kg to 20 kg. Max grouped load 500 kg.  2. Other loads within the overall listed range may also be used.	S
END				