


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION</p> <p><b>0467</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Devon Metrology Laboratory</h3> <p>Issue No: 011    Issue date: 26 May 2011</p>	
	<p><b>Trading Standards Service</b> County Hall Topsham Road Exeter EX2 4QH</p>	<p><b>Contact: Mrs Janet Rudd</b> Tel: +44 (0)1392 383316 / 381381 Fax: +44 (0)1392 382732 E-Mail: <a href="mailto:calibration@devon.gov.uk">calibration@devon.gov.uk</a> Website: <a href="http://devon.gov.uk/calibration">devon.gov.uk/calibration</a></p>
<p><b>Calibration performed at the above address only</b></p>		

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
MASS	Nominal value (g)	(mg)	
Weights and artefacts	25 000	75	<p>1. Calibrations can be given in other units as required.</p> <p>2. Intermediate values can be calibrated with an uncertainty interpolated from the next higher and lower values in the table above.</p>
	20 000	60	
	10 000	30	
	5 000	15	
	2 000	2.0	
	1 000	1.0	
	500	0.50	
	200	0.10	
	100	0.05	
	50	0.033	
	20	0.027	
	10	0.020	
	5	0.017	
	2	0.013	
	1	0.010	
	0.5	0.0080	
	0.2	0.0070	
	0.1	0.0050	
	0.05	0.0040	
	0.02	0.0030	
0.01	0.0030		
0.005	0.0020		
0.002	0.0020		
0.001	0.0020		
END			