

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>0474</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Hertfordshire Metrology Laboratory</h3> <p>Issue No:014 Issue date: 02 August 2010</p>	
	<p>Apsley 2 Brindley Way Hemel Hempstead Herts HP3 9BF</p>	<p>Contact: Mr A Burrows Tel+44 (0)1442 454505 Fax: +44 (0)1442 454501 E-Mail: allan.burrows@hertscc.gov.uk Website:</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks
MASS	Nominal value (g)	(mg)	NOTE
Weights and artefacts	25 000	12.5	<ol style="list-style-type: none"> 1. Calibrations can be given in other units as required. 2. Intermediate values can be calibrated to an uncertainty interpolated from the next higher and lower values in the table.
	20 000	10	
	10 000	5.0	
	5 000	2.5	
	2 000	1.0	
	1 000	0.50	
	500	0.25	
	200	0.10	
	100	0.050	
	50	0.033	
	20	0.027	
	10	0.020	
	5	0.017	
	2	0.013	
	1	0.010	
	0.5	0.0080	
	0.2	0.0070	
	0.1	0.0050	
	0.05	0.0040	
	0.02	0.0030	
0.01	0.0030		
0.005	0.0020		
0.002	0.0020		
0.001	0.0020		
END			