


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;"><b>0477</b></p> <p style="text-align: center;">Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3 style="margin: 0;">Dstl Radiation Protection Services</h3> <p style="margin: 0;">Issue No: 013    Issue date: 30 June 2010</p>	
	<p>At the Institute of Naval Medicine Crescent Road Alverstoke Gosport Hampshire PO12 2DL</p>	<p>Contact: Mr A French Tel: +44 (0)239 276 8010 Fax: +44 (0)239 276 8265 E-Mail: <a href="mailto:afrench@dstl.gov.uk">afrench@dstl.gov.uk</a> Website:</p>
<p><b>Calibration performed at the above address only</b></p>		

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AIR KERMA RATE	<sup>137</sup> Cs: 2.0 $\mu\text{Gyh}^{-1}$ to 1.0 $\text{Gyh}^{-1}$ <sup>60</sup> Co: 35 $\mu\text{Gyh}^{-1}$ to 20 $\text{mGyh}^{-1}$ <sup>241</sup> Am: 5 $\mu\text{Gyh}^{-1}$ to 200 $\mu\text{Gyh}^{-1}$	<p>3.0 %</p> <p>3.0 %</p> <p>5.0 %</p>	<p>Air Kerma and Ambient dose Equivalent rate fields generated in accordance with ISO 4037-1 (1966)</p> <p>Dosimetry and conversion coefficients conform to ISO 4037-2 (1997) and ISO 4037-3 (1999)</p>
AMBIENT DOSE EQUIVALENT RATE	<sup>137</sup> Cs: 2.5 $\mu\text{Svh}^{-1}$ to 1.2 $\text{Svh}^{-1}$ <sup>60</sup> Co: 40 $\mu\text{Svh}^{-1}$ to 23 $\text{mSvh}^{-1}$ <sup>241</sup> Am: 9 $\mu\text{Svh}^{-1}$ to 350 $\mu\text{Svh}^{-1}$	<p>3.0 %</p> <p>3.0 %</p> <p>5.0 %</p>	
PERSONAL DOSE EQUIVALENT:			
PERFORMANCE TESTING OF DOSIMETRY SERVICES FOR EXTERNAL, WHOLE BODY GAMMA RADIATION;	<sup>137</sup> Cs: to HSE Protocols	Uncertainties for <sup>137</sup> Cs Air Kerma: 3.0 %	
PERFORMANCE TESTING OF DOSIMETRY SERVICES FOR EXTREMITY/SKIN GAMMA RADIATION;	As above	As above	
PERFORMANCE TESTING OF ACCIDENT DOSIMETRY SERVICES - WHOLE BODY GAMMA RADIATION;	As above	As above	
ROUTINE IRRADIATION OF PERSONAL DOSEMETERS	As above	As above	



0477  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Dstl Radiation Protection Services**  
Issue No: 013    Issue date: 30 June 2010

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
SURFACE CONTAMINATION MONITOR RESPONSE:  ALPHA ( $\alpha$ ) CONTAMINATION BETA ( $\beta$ ) CONTAMINATION GAMMA ( $\gamma$ ) AND X-RAY CONTAMINATION  Statutory tests, including: Test before first use; Periodic tests	Alpha-emitting nuclides: $^{241}\text{Am}$ , $^{238}\text{Pu}$ , $^{238}\text{U}$ , $^{230}\text{Th}$ Beta-emitting nuclides: $^{90}\text{Sr}$ , $^{90}\text{Y}$ , $^{147}\text{Pm}$ , $^{36}\text{Cl}$ , $^{14}\text{C}$ , $^{60}\text{Co}$ , $^{137}\text{Cs}$ , $^{63}\text{Ni}$ Photon-emitting nuclides: $^{55}\text{Fe}$ , $^{238}\text{Pu}$ , $^{129}\text{I}$ , $^{241}\text{Am}$ , $^{57}\text{Co}$ , $^{137}\text{Cs}$ , $^{60}\text{Co}$	5.0% to 15 % depending upon monitor type	
END			