

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="margin: 0;">Accredited to ISO/IEC 17025:2005</p>	<p>LMS Ltd</p> <p>Issue No: 014 Issue date: 18 August 2011</p>	
	<p>The Modern Forge Riverhead Sevenoaks Kent TN13 2EL</p>	<p>Contact: Mr M D Jeal Tel: +44 (0)1732 451866 Fax: +44 (0)1732 450127 E-Mail: sales@lms.ltd.uk Website: www.lms.ltd.uk</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Activities performed on site*:

Location details	Activity	Location code
Customer's Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of LMS)	Humidity (%rh Enclosures) Rotational Speed (Centrifuges) Temperature (Sterilizers, etc, and Thermal Products)	S



0503
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

LMS Ltd
Issue No: 014 **Issue date:** 18 August 2011

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
Temperature controlled autoclaves, media preparators, incubators, sterilizers, digestion blocks, freezers, refrigerators, ovens, environmental chambers, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)			A list of individual approved engineers is held by the laboratory and by UKAS	
TEMPERATURE	- 10 °C to 105 °C	0.26 °C using resistance sensors	Single and multipoint monitoring probes Time dependent temperature profiling	S
	- 70 °C to 400 °C	1.9 °C		
	400 °C to 800 °C	4.0 °C		
	800 °C to 1000 °C	6.1 °C using thermocouple sensors		
RELATIVE HUMIDITY	Environmental chambers 25 %rh to 90 %rh (20 °C to 40 °C)	2.0 %rh		S
ROTATIONAL SPEED	Centrifuges 1000 rpm to 20 000 rpm	4.1 rpm		S
TIME INTERVAL	Timers 10 s to 1 hr	1.3 s		S
END				