


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>0548</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>3M Health Care Limited</b>	
	Issue No: 019    Issue date: 05 December 2011	
	<b>3M Service Centre</b> <b>5 Seymour Court</b> <b>Manor Park</b> <b>Runcorn</b> <b>Cheshire</b> <b>WA7 1UZ</b>	<b>Contact: Mr R Norton</b> <b>Tel: +44 (0)1928 532590</b> <b>Fax: +44 (0)1928 597769</b> <b>E-Mail: 3mruncorn@mmm.com</b> <b>Website: www.3m.com</b>
<b>Calibration performed by the Organisations at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> 3M Service Centre 5 Seymour Court Manor Park Runcorn Cheshire WA7 1UZ	<b>Local contact</b> Mr Rick Norton  Tel: +44 (0)1928 532590 Fax: +44 (0)1928 597769 Email: 3mruncorn@mmm.com Website: www.3m.com	Electrical Flow Pressure Temperature	P

#### Site activities performed away from the locations listed above:

Location details		Activity	Location code
Customer Premises - Any	Contact as above	Electrical Pressure Temperature	S



0548

Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**3M Health Care Limited**

**Issue No: 019 Issue date: 05 December 2011**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks	Location Code
<b>ELECTRICAL</b>				
TIME INTERVAL				
Voltage DC	9 V to 11 V	64 mV		P
Timers	10 s to 60 min	0.17 s		P & S
<b>FLOW</b>				
Air Flow (Air Samples)	15 l/min to 30 l/min 28.3 l/min	2.1 % 0.83 %		P
<b>PRESSURE</b>				
Gas pressure (gauge)	- 100 kPa to 400 kPa	2.7 kPa		P & S
<b>TEMPERATURE</b>				
Temperature indicators, controllers and recorders all with temperature sensors.	0 °C to 130 °C	0.22 °C	Plus resolution of the Instrument	P & S
Temperature controlled autoclaves, media preparators, chambers, environmental cabinets and ovens and similar apparatus.	130 °C to 150 °C	0.35 °C		
END				