


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

|  |  |  |
|--|--|--|
|  <p><b>UKAS</b><br/>CALIBRATION</p> <p><b>0548</b></p> <p>Accredited to<br/><b>ISO/IEC 17025:2005</b></p> | <p><b>3M Health Care Limited</b></p> <p>Issue No:016    Issue date: 17 July 2009</p>                     |  |
|  | <p><b>3M Service Centre</b><br/>11 Seymour Court<br/>Manor Park<br/>Runcorn<br/>Cheshire<br/>WA7 1SY</p> | <p><b>Contact: Mr R Norton</b><br/>Tel: +44 (0)1928 532590<br/>Fax: +44 (0)1928 597769<br/>E-Mail: 3mruncorn@mmm.com<br/>Website: www.3m.com</p> |
| <p><b>Calibration performed by the Organisations at the locations specified below</b></p>  |  |  |

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

| Location details   |  | Activity  | Location code |
|--|--|---|---------------|
| <p><b>Address</b><br/>3M Service Centre<br/>11 Seymour Court<br/>Manor Park<br/>Runcorn<br/>Cheshire<br/>WA7 1SY</p> | <p><b>Local contact</b><br/>Mr Rick Norton<br/><br/>Tel: +44 (0)1928 532590<br/>Fax: +44 (0)1928 597769<br/>Email: 3mruncorn@mmm.com<br/>Website: www.3m.com</p> | <p>Electrical<br/>Flow<br/>Pressure<br/>Temperature</p> | P             |

#### Site activities performed away from the locations listed above:

| Location details        |                  | Activity                                       | Location code |
|-------------------------|------------------|--|---------------|
| Customer Premises - Any | Contact as above | <p>Electrical<br/>Pressure<br/>Temperature</p> | S             |



0548  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**3M Health Care Limited**  
**Issue No: 016 Issue date: 17 July 2009**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Measured Quantity<br>Instrument or Gauge   | Range                | Best<br>Measurement<br>Capability<br>Expressed as an<br>Expanded<br>Uncertainty ( $k=2$ ) | Remarks                              | Location<br>Code |
|--|----------------------|---|--------------------------------------|------------------|
| <b>ELECTRICAL</b>  |                      |   |                                      |                  |
| TIME INTERVAL  |                      |   |                                      |                  |
| Voltage DC   | 9 V to 11 V          | 64 mV   |                                      | P                |
| Timers   | 10 s to 60 min       | 0.17 s  |                                      | P & S            |
| <b>FLOW</b>  |                      |   |                                      |                  |
| Air Flow (Air Samples)   | 20 l/min to 35 l/min | 0.55 %  |                                      | P                |
| <b>PRESSURE</b>  |                      |   |                                      |                  |
| Gas pressure (gauge)   | -100 kPa to 400 kPa  | 3.1 kPa   |                                      | P & S            |
| <b>TEMPERATURE</b>   |                      |   |                                      |                  |
| Temperature indicators,<br>controllers and recorders all<br>with temperature sensors.<br>Temperature controlled<br>autoclaves, media preparators,<br>chambers, environmental<br>cabinets and ovens and similar<br>apparatus. | 0 °C to 130 °C       | 0.29 °C   | Plus resolution of the<br>Instrument | P & S            |
|  | 130 °C to 150 °C     | 0.35 °C   |                                      |                  |
| END  |                      |   |                                      |                  |