


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>0564</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>TC Ltd</p> <p>Issue No: 022 Issue date: 02 April 2012</p>	
	<p>Brimington Road North Whittington Moor Chesterfield Derbyshire S41 9BE</p>	<p>Contact: Mr L R Walker Tel: +44 (0)1246 454672 Fax: +44 (0)1895 273540 E-Mail: laboratory@tc.co.uk Website: www.tc.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
TEMPERATURE			
Resistance thermometers	Liquid nitrogen (-196 °C) Triple point of water (0.01 °C) -100 °C to -30 °C - 30 °C to + 20 °C 20 °C to 90 °C 90 °C to 200 °C 200 °C to 400 °C 400 °C to 620 °C	0.20 °C 0.050 °C 0.40 °C 0.20 °C 0.20 °C 0.20 °C 0.40 °C 0.40 °C	2 wire, 3 wire and 4 wire resistance thermometers
Thermocouples - noble metal	20 °C to 200 °C 200 °C to 400 °C 400 °C to 620 °C 620 °C to 1100 °C 1100 °C to 1330 °C 1330 °C to 1590 °C	1.0 °C 1.0 °C 1.0 °C 1.0 °C 1.0 °C 1.7 °C	Extension and Compensating Cables
Thermocouples – base metal	Liquid nitrogen (-196 °C) -100 °C to -30 °C - 30 °C to + 200 °C 200 °C to 400 °C 400 °C to 620 °C 620 °C to 1100 °C 1100 °C to 1330 °C	0.40 °C 0.40 °C 0.30 °C 0.40 °C 0.40 °C 0.80 °C 0.90 °C	
Temperature indicators and transmitters with probes	Ranges as for sensor type	As for sensor type	Including instruments with electrical outputs
Electrical calibration of temperature simulators for the following sensors			
Resistance sensors (PT 100)	-200 °C to +800 °C	0.05 °C	2 wire, 3 wire and 4 wire resistance sensors
Noble metal thermocouples	-20 °C to +1750 °C	0.80 °C	Including cold junction compensation
Base metal thermocouples	-200 °C to +1372 °C	0.40 °C	Including cold junction compensation



0564
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

TC Ltd
Issue No: 022 Issue date: 02 April 2012

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
Electrical calibration of temperature indicators, controllers and recorders for the following sensors			
Resistance sensors (PT 100)	-200 °C to +800 °C	0.05 °C	2 wire, 3 wire and 4 wire resistance sensors
Noble metal thermocouples	-20 °C to +1750 °C	0.80 °C	Including cold junction compensation
Base metal thermocouples	-200 °C to +1372 °C	0.40 °C	Including cold junction compensation
END			