

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>0580</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Furness Controls Limited</h3> <p>Issue No: 028 Issue date: 10 February 2012</p>	
	<p>Beeching Road Bexhill East Sussex TN39 3LJ</p>	<p>Contact: Mr D Walker Tel: +44 (0)1424-730316 Fax: +44 (0)1424-730317 E-Mail: calibration@furness-controls.com Website: www.furness-controls.com</p>
<p>Calibration performed by the Organisation at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
<p>Address Beeching Road Bexhill East Sussex TN39 3LJ</p> <p>Local contact Doug Walker</p>	<p>Flow calibration Pressure calibration Electrical Calibration</p>	Perm
<p>Techniparc 3 rue Boole 91240 St. Michel sur Orge France</p> <p>Jean-Philippe Noblet Tel.+33 1 69460020</p>	<p>Flow calibration Pressure calibration Electrical Calibration</p>	France and site
<p>Karl Arnold Strasse 12 D-47877 Willich Munchheide (2) Germany</p> <p>Karsten Bartsch Tel. +49 21 54 49 96 80</p>	<p>Flow calibration Pressure calibration Electrical Calibration</p>	Germany and site

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>The customer's site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer</p> <p>Site contact: Sarah Hedge 4 The Pavilions Amber Close Tamworth Staffordshire B77 4RP Tel: +44 (0)1827 59950</p>	<p>Flow calibration Pressure calibration</p>	Site



0580
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Furness Controls Limited
Issue No: 028 Issue date: 10 February 2012

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
PRESSURE				
<u>Gas pressure (gauge)</u> Calibration of pressure indicating instruments and gauges Pressure equivalent calibration of Furness controls FRS 4's and other dead weight testers	-100 kPa to - 10 kPa - 10 kPa to 0 Pa 0 Pa to 3 kPa 3 kPa to 12 kPa 12 kPa to 30 kPa 30 kPa to 40 kPa 40 kPa to 4 MPa	0.010 % + 30 Pa 0.010 % + 0.30 Pa 0.010 % + 0.030 Pa 0.010 % + 0.30 Pa 0.010 % + 1.0 Pa 0.010 % + 2.0 Pa 0.010 %	Calibrations of devices with an electrical output may be undertaken	Perm
<u>Gas pressure (absolute)</u> Calibration of pressure indicating instruments and gauges	0 Pa to 130 kPa	0.010 % + 10 Pa	Absolute instruments can be calibrated over the gauge pressure range using barometric pressure measurement with an additional uncertainty of 15 Pa	Perm
<u>Gas pressure (gauge)</u> Calibration of pressure indicating instruments and gauges	- 100 kPa to 0 Pa 0 Pa to 220 Pa 220 Pa to 2200 Pa 2.2 kPa to 22 kPa 22 kPa to 100 kPa 100 kPa to 400 kPa 400 kPa to 1.6 MPa 1.6 MPa to 4 MPa	0.25 % + 100 Pa 0.30 % + 0.10 Pa 0.30 % + 0.60 Pa 0.30 % + 6.0 Pa 0.25 % + 50 Pa 0.25 % + 200 Pa 0.25 % + 800 Pa 0.25 % + 2.0 kPa		Site
<u>Gas pressure (absolute)</u> Calibration of pressure indicating instruments and gauges	0 Pa to 160 kPa	0.25 % + 100 Pa		Site
FLOW				
Flow Rate - Gas Volume	0.02 ml/min to 2ml/min 2 ml/min to 2000 l/min	0.75 % 0.50 %	Calibration medium Air. Calibrations up to 10 l/min can be undertaken on Nitrogen.	Perm
Gas - Volume Passed	180 l to 300 l (at flow rates of 2 l/min to 2000 l/min)	0.36 %		Perm
	0.02 ml to 6000 l (at flow rates of 0.02 ml/min to 2ml/min 2 ml/min to 2000 l/min)	0.75 % 0.50 %		Perm



0580
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Furness Controls Limited
Issue No: 028 Issue date: 10 February 2012

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
FLOW Gas volume flow rate	0.2 ml/min to 40 ml/min 40 ml/min to 2000 l/min	1.2 % 1.0 %	Calibration medium Air.	Site
ELECTRICAL DC Voltage Measurement	0 V to 10 V 10 V to 55 V	30 ppm + 20 μ V 30 ppm + 200 μ V		Perm France Germany
DC Current Measurement	0 A to 10 mA 10 mA to 100 mA	100 ppm + 0.30 μ A 100 ppm + 3.0 μ A		Perm France Germany
French Capability			Calibrations of devices with an electrical output may be undertaken	France
FLOW Flow Rate - Gas Volume	0.2 ml/min to 40 ml/min 40 ml/min to 400 l/min	1.2 % 1.0 %	Calibration medium Air	
PRESSURE <u>Gas pressure (gauge)</u> Calibration of pressure indicating instruments and gauges Furness Controls FRS 4s, pressure equivalent	- 100 kPa to 0 Pa 0 Pa to 3.2 kPa 3.2 kPa to 12 kPa 12 kPa to 30 kPa 30 kPa to 40 kPa 40 kPa to 100 kPa 100 kPa to 1.6 MPa 1.6 MPa to 4 MPa	0.10 % + 100 Pa 0.012 % + 0.035 Pa 0.012 % + 0.40 Pa 0.012 % + 1.0 Pa 0.012 % + 2.0 Pa 0.10 % + 20 Pa 0.10 % + 150 Pa 0.10 % + 200 Pa		
<u>Gas pressure (absolute)</u> Calibration of pressure indicating instruments and gauges	1 kPa to 160 kPa 160 kPa to 200 kPa	0.10 % + 100 Pa 0.10 % + 200 Pa		
German Capability			Calibrations of devices with an electrical output may be undertaken	Germany
FLOW Flow Rate - Gas Volume	4 ml/min to 40 ml/min 40 ml/min to 400 l/min	1.2 % 1.0 %	Calibration medium Air	



0580
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Furness Controls Limited
Issue No: 028 Issue date: 10 February 2012

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
<p>PRESSURE</p> <p><u>Gas pressure (gauge)</u></p> <p>Calibration of pressure indicating instruments and gauges Furness controls FRS 4's. pressure equivalent</p> <p><u>Gas pressure (absolute)</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>- 100 kPa to 0 Pa 0 Pa to 3.2 kPa 3.2 kPa to 12 kPa 12 kPa to 30 kPa 30 kPa to 40 kPa 40 kPa to 400 kPa 400 kPa to 1.6 MPa</p> <p>1 kPa to 120 kPa</p>	<p>0.15 % + 100 Pa 0.013 % + 0.035 Pa 0.012 % + 0.50 Pa 0.012 % + 1.0 Pa 0.012 % + 2.0 Pa 0.15 % + 50 Pa 0.15 % + 150 Pa</p> <p>0.15 % + 100 Pa</p>		Germany
END				