


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

| | | |
|--|---|---|
|  <p>UKAS CALIBRATION</p> <p>0582</p> <p>Accredited to ISO/IEC 17025:2005</p> | <p>S Brannan and Sons Ltd</p> <p>Issue No: 013 Issue date: 15 June 2010</p> | |
| | <p>Leconfield Industrial Estate Cleator Moor Cumbria CA25 5QE</p> | <p>Contact: Mr R Hellon Tel: +44 (0)1946 816600 Fax: +44 (0)1946 816625 E-Mail: richard.hellon@brannan.co.uk Website: http://www.brannan.co.uk</p> |
| <p>Calibration performed at the above address only</p> | | |

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$) | Remarks |
|--|--|---|---------|
| TEMPERATURE | | | |
| Liquid-in-glass thermometers and dial analogue thermometers | - 60 °C to - 40 °C - 40 °C to - 10 °C - 10 °C to + 0.01 °C + 0.01 °C (water triple point) | 0.23 °C 0.040 °C 0.040 °C 0.010 °C | |
| | 0.01 °C to 100 °C 100 °C to 156 °C 156 °C to 250 °C | 0.040 °C 0.060 °C 0.070 °C | |
| Electronic thermometers with indicators | As above | As above | |
| <p>END</p> | | | |