


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 UKAS CALIBRATION 0594 Accredited to ISO/IEC 17025:2005	Astell Scientific Ltd	
	Issue No: 013 Issue date: 09 July 2010	
	Powerscroft Road Sidcup Kent DA14 5DT	Contact: Mr B Bunning Tel: +44 (0)20 8309 2039 Fax: +44 (0)20 8309 2009 E-Mail: bbunning@astell.com Website: www.astell.com
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above*:

Location details	Activity	Location code
Customer Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of Astell)	Temperature (Sterilizers etc, and Thermal Products)	S



0594
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Astell Scientific Ltd
Issue No: 013 Issue date: 09 July 2010

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, environmental cabinets, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges).	50 °C to 150 °C	0.60 °C	A list of individual approved engineers is held by the laboratory and by UKAS. Single and multipoint monitoring probes. Time dependent temperature profiling.	S
TIME INTERVAL Timers	60 s to 5 hr	7.0 s		S
END				