


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0602 Accredited to ISO/IEC 17025:2005	Priorclave Ltd	
	Issue No: 010 Issue date: 09 July 2010	
	129/131 Nathan Way West Thamesmead Business Park Woolwich London SE28 0AB	Contact: Mr P R Melling Tel: +44 (0)20 8316 6620 Fax: +44 (0)20 8855 9374 E-Mail: petermelling@priorclave.co.uk Website: www.priorclave.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above*:

Location details	Activity	Location code
Customer Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of Priorclave)	Temperature (Sterilizers etc, and Thermal Products)	S



0602
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Priorclave Ltd
Issue No: 010 Issue date: 09 July 2010

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, refrigerators, freezers, environmental cabinets, chambers, and rooms (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	- 40 °C to + 200 °C	0.50 °C*	A list of individual approved engineers is held by the laboratory and by UKAS Single and multipoint monitoring probes Time dependent temperature profiling	S
TIME INTERVAL Timers	10 s to 2 hr	2.0 s	* plus resolution of indicator/recorder	S
END				