


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0605 Accredited to ISO/IEC 17025:2005	Status Metrology Solutions Ltd	
	Issue No:011 Issue date: 04 October 2010	
	Measurement House Lenton Street Sandiacre Nottingham NG10 5DX	Contact: Mr D Wright Tel: +44 (0)115 939 2228 Fax: +44 (0)115 939 3355 E-Mail: info@statusmetrology.com Website: www.statusmetrology.com
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Measurement House Lenton Street Sandiacre Nottingham NG10 5DX	Local contact Mr D Wright	Dimensional A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
At customers premises	Mr D Wright	Dimensional B



0605
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Status Metrology Solutions Ltd
Issue No:011 Issue date: 04 October 2010

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
Performance verification of co-ordinate measuring machines	As ISO 10360-2:2001 0 mm to 1800 mm (longest diagonal using end standards)	(1.3 + 1.0 per metre) μm		A & B
END				