


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0617 Accredited to ISO/IEC 17025:2005	LTE Scientific Ltd	
	Issue No: 011 Issue date: 16 December 2011	
	Greenbridge Lane Greenfield Oldham Lancashire OL3 7EN	Contact: Kim Cooper Tel: +44 (0)1457 876221 Fax: +44 (0)1457 870131 E-Mail: k.cooper@lte-scientific.co.uk Website: www.lte-scientific.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of LTE)	Temperature (Sterilizers etc, and Thermal Products) Pressure Time	S



0617
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

LTE Scientific Ltd
Issue No: 011 Issue date: 16 December 2011

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, environmental cabinets, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges).	0 °C to 140 °C 140 °C to 250 °C	0.63°C 1.2 °C	A list of individual approved engineers is held by the laboratory and by UKAS. Single and multipoint monitoring probes. Time dependent temperature profiling.	S
TIME INTERVAL Timers	30 s to 3600 s	4.0 s		S
PRESSURE Gas pressure (gauge)	- 0.9 kPa to 300 kPa	6 kPa		S
END				