

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>0640</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Hanna Instruments Ltd</h3> <p>Issue No: 009 Issue date: 01 December 2010</p>	
	<p>Eden Way Pages Industrial Park Leighton Buzzard Bedfordshire LU7 4AD</p>	<p>Contact: Mr M Shenton Tel: +44 (0)1525 850855 Fax: +44 (0)1525 853668 E-Mail: martin.shenton@hannainst.co.uk Website: www.hannainst.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
TEMPERATURE			
Temperature sensors with indicators			Calibration by comparison
Resistance sensors	- 20 °C to + 90 °C 90 °C to 180 °C	0.030 °C 0.18 °C	
Thermocouple sensors	- 20 °C to + 90 °C 90 °C to 180 °C	0.13 °C 0.28 °C	
END			