


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>0669</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Novatron</h3> <p>Issue No: 020 Issue date: 15 December 2010</p>	
	<p>Novatron Scientific Ltd Novatron House 46 London Road Horsham West Sussex RH12 1AY</p>	<p>Contact: Mr S H Long Tel: +44 (0)1403 754416 Fax: +44 (0)1403 754480 E-Mail: simonhlong@btopenworld.com Website: www.novatron.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
<p>RELATIVE HUMIDITY</p> <p>Relative humidity meters</p>	<p>6 %rh to 50 %rh 50 %rh to 95 %rh 95 %rh to 98 %rh <i>for the temperature range 20 °C to 25 °C</i></p>	<p>1.6 %rh 2.2 %rh 2.7 %rh</p>	<p>Calibrations of devices with an electrical output may be undertaken.</p>
<p>Saturated salt capsules</p>	<p>6 %rh to 11 %rh 33 %rh 53 %rh 75 %rh 90 %rh 98 %rh</p>	<p>1.0 %rh 1.2 %rh 1.2 %rh 1.5 %rh 1.7 %rh 2.2 %rh</p>	<p>The accreditation covers other humidity units directly related to relative humidity, eg, equilibrium relative humidity (ERH).</p>
<p>TEMPERATURE</p> <p>Temperature sensors with indicators (measurements made in air)</p>	<p>20 °C to 25 °C</p>	<p>0.20 °C</p>	<p>Calibrations of devices with an electrical output may be undertaken.</p>
<p>END</p>			