


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>0670</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Measure-Rite Limited</h3> <p>Issue No:007    Issue date: 06 July 2010</p>	
	<p>Tuckey's Farm Cathiron Lane Nr. Harborough Magna Rugby Warwickshire CV23 0JH</p>	<p>Contact: Mr Gary Robertson Tel: +44 (0) 1788 833360 Fax: +44 (0) 1788 833358 E-Mail: sales@measure-rite.com Website: www.measure-rite.com</p>
<p><b>Calibration performed by the Organisations at the locations specified below</b></p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b></p> <p><b>Local contact</b></p>		
<p><b>Address</b> Tuckey's Farm Cathiron Lane Nr. Harborough Magna Rugby Warwickshire CV23 0JH</p> <p><b>Local contact</b> Mr Gary Robertson</p>	Dimensional	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p><b>Address</b></p> <p><b>Local contact</b></p>		
<p>At customers premises</p> <p>Mr Gary Robertson</p>	Dimensional	B



0670  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Measure-Rite Limited**  
**Issue No: 007 Issue date: 06 July 2010**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
FORM			NOTES	
Tables, departure from horizontal plane. (Excluding calibrations to BS 817:1988)	300 to 20 000	0.060 + (0.01 x diagonal in m) mm (See Note 1)	1. The uncertainty quoted is for the departure of the table surface from a true horizontal plane.  2. All calibrations can be given in inch units.	B
MEASURING INSTRUMENTS AND MACHINES				
Tool Pre-setters	100 to 300	12		B
Measuring Machines, 3 Axis (Excluding Co-ordinate Measuring Machines to ISO 10360)	300 to 20 000	Linear - 0.10 mm per metre Squareness - 0.050 mm per metre		B
END				