


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>0685</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Mectronics Marine</b>	
	Issue No: 009    Issue date: 16 April 2007	
	7 Ware Close Kingsteignton Newton Abbot Devon TQ12 3DS	Contact: Mr M E Chaney Tel: +44 (0)1626 334453 Fax: +44 (0)1626 361761 E-Mail: <a href="mailto:calibration@mectronics.co.uk">calibration@mectronics.co.uk</a> Website: <a href="http://www.mectronics.co.uk">www.mectronics.co.uk</a>
Calibration performed at the above address only		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
ANTENNA FACTOR  Loop Antennas	5 Hz to 1 kHz 1 kHz to 3 MHz 3 MHz to 30 MHz	0.8 dB 0.6 dB 1.1 dB	Up to 1.2 m diameter IEEE 291-1991
END			