


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 UKAS CALIBRATION 0711 Accredited to ISO/IEC 17025:2005	IncoTest	
	Issue No: 004	Issue date: 20 September 2011
	Special Metals Wiggin Limited Holmer Road Hereford HR4 9SL	Contact: Dr J R Silk Tel: +44 (0)1432 352230 Fax: +44 (0)1432 353545 E-Mail: jonathon.silk@incotest.co.uk Website: www.incotest.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Special Metals Wiggin Limited Holmer Road Hereford HR4 9SL	Local contact Dr J R Silk	Temperature	A



0711
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

IncoTest

Issue No: 004 Issue date: 20 September 2011

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
TEMPERATURE				
Platinum thermocouples	100 °C to 600 °C 600 °C to 1100 °C 1100 °C to 1300 °C	1.4 °C to 1.1 °C 1.1 °C 2.5 °C		A
Base Metal thermocouples	100 °C to 1100 °C 1100 °C to 1300 °C	1.2 °C 2.6 °C		A
END				