


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION</p> <p><b>0797</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Lancashire County Council</h3> <p><b>Issue No: 009    Issue date: 19 August 2011</b></p>	
	<p><b>Trading Standards</b> 58-60 Guildhall Street Preston PR1 3NU</p>	<p><b>Contact: Daniel Maxim</b> <b>Tel: +44 (0) 1772 533564</b> <b>Fax: +44 (0) 1772 533648</b> <b>E-Mail: calibration@lancashire.gov.uk</b> <b>Website: www.weightsandmeasures.co.uk</b></p>
<p><b>Calibration performed at the above address only</b></p>		

#### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
MASS	Nominal value (g)	(mg)	NOTES
Weights and artefacts	25 000	12.5	<p>1. Calibrations can be given in other units as required.</p> <p>2. Intermediate values can be calibrated to an uncertainty interpolated from the next higher and lower values in the table.</p>
	20 000	10	
	10 000	5.0	
	5 000	2.5	
	2 000	1.0	
	1 000	0.50	
	500	0.25	
	200	0.10	
	100	0.05	
	50	0.033	
	20	0.027	
	10	0.020	
	5	0.017	
	2	0.013	
	1	0.010	
	0.5	0.0080	
	0.2	0.0070	
	0.1	0.0050	
	0.05	0.0040	
	0.02	0.0030	
	0.01	0.0030	
	0.005	0.0020	
	0.002	0.0020	
	0.001	0.0020	
END			