


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 0812 Accredited to ISO/IEC 17025:2005	KROHNE Ltd	
	Issue No: 009 Issue date: 10 June 2010	
	Rutherford Drive Park Farm South Industrial Estate Wellingborough Northants NN8 6AE	Contact: Mr G Howe Tel: +44 (0)1933 408500 Fax: +44 (0)1933 408501 E-Mail: g.howe@krohne.com Website: www.krohne.com
Calibration performed at the above address only		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
FLOW Mass Flow (Water) Quantity passed	10 kg to 5000 kg (At flow rates of 95 kg/h to 850 000 kg/h)	0.035 %	Suitable for Coriolis meters (own manufacture)
Volume flow (Water) Quantity passed	10 l to 5000 l (At flow rates of 95 kg/h to 850 000 kg/h)	0.038 %	
END			