


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

| | | |
|--|---|--|
|  0814 Accredited to ISO/IEC 17025:2005 | Medical Physics International Ltd Issue No: 009 Issue date: 05 February 2012 | |
| | Unit 9-11 Bolton Technology Exchange Spa Road Bolton BL1 4AY | Contact: Paul Perkins Tel: +44 (0) 1204 555900 Fax: +44 (0) 1204 555909 E-Mail: service@medphysics.co.uk Website: www.medphysics.co.uk |

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above*:

| Location details | Activity | Location code |
|--|--------------------------------|---------------|
| Customer's Premises, eg, Hospitals, Laboratories, and Manufacturing Plants (*including those of MPI) | Temperature (Thermal Products) | S |



0814
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Medical Physics International Ltd
Issue No: 009 Issue date: 05 February 2012

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

| Measured Quantity Instrument or Gauge | Range | Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$) | Remarks | Location Code |
|--|-------------------|--|---|------------------|
| TEMPERATURE Temperature controlled incubators, ovens, environmental chambers, refrigerators, and freezers (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges) | - 30 °C to 140 °C | 0.50 °C | A list of individual approved engineers is held by the laboratory and by UKAS Using 'Venus blocks' Single and multipoint monitoring probes Time dependent temperature profiling | S |
| END | | | | |