


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 UKAS CALIBRATION 4057 Accredited to ISO/IEC 17025:2005	Interhaze Global Solutions Ltd (trading as) Iso-Tech Calibration Services Issue No: 005 Issue date: 08 November 2011	
	Unit 3 Riverside Park East Service Way Raynesway Derby DE21 7RW	Contact: Mr Richard Gore Tel: +44 (0)1332 664245 Fax: +44 (0)1332 668399 E-Mail: enquiries@iso-tech.org Website: www.iso-tech.org

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers premises	Dimensional	A



4057
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Interhaze Global Solutions Ltd
(trading as) Iso-Tech Calibration Services
Issue No: 005 Issue date: 08 November 2011

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED				
MEASURING INSTRUMENTS AND MACHINES				
Performance verification of co-ordinate measuring machines	As ISO 10360-2:2002 0 to 2743 (longest diagonal using end standards)	1.5 + (1.4 x length in m)		A
Performance verification of co-ordinate measuring machines	As ISO 10360-2:2009 and ISO 10360-5:2010 0 to 2743 (longest diagonal using end standards)	1.5 + (1.4 x length in m)		A
Performance verification of non-cartesian co-ordinate measuring machines	In house method based on ISO 10360-2:2009 0 to 2743 (maximum diameter using end standards)	2.2 + (2.2 x length in m)		A
END				