


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>4185</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>Paul MacArthur Limited trading as SCCS</b>	
	Issue No: 003    Issue date: 18 August 2010	
	HQI Building Phoenix Park Eaton Socon Cambridgeshire PE19 8EP	Contact: Mr P Kelly Tel: +44 (0)1634 296809 Fax: +44 (0)1634 723055 E-Mail: peter.kelly@sccssurvey.co.uk Website: www.sccssurvey.co.uk
Calibration performed by the Organisations at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Unit 4 Sir Thomas Langley Road Medway City Estate Rochester Kent ME2 4DP	Electronic Distance Measurement	A



4185  
Accredited to  
ISO/IEC 17025:2005

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

**Paul MacArthur Limited trading as SCCS**  
**Issue No: 003 Issue date: 18 August 2010**

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks	Location Code
BASE-LINE LENGTH	0 to 17 m	0.42 mm	1. The base line is located at the Rochester Laboratory	A
ELECTRONIC DISTANCE MEASURING EQUIPMENT	0 to 17 m	0.50 mm	2. The EDM equipment will be calibrated against the base line located at the Rochester Laboratory	A
END				