


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 UKAS CALIBRATION 4296 Accredited to ISO/IEC 17025:2005	STAR LAB (UK) Ltd Issue No: 004 Issue date: 19 March 2012	
	4 Tanners Drive Blakelands Milton Keynes MK14 5NA	Contact: Mr Andy Hall Tel: +44 (0)1908 283800 Fax: +44 (0)1908 283802 E-Mail: AndyHall@starlab.co.uk Website: www.starlab.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 4 Tanners Drive Blakelands Milton Keynes MK14 5NA Local contact Andy Hall	Volume	P

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customers' sites or premises The customer's sites or premises must be suitable for the nature of the particular calibrations undertaken and will be subject of contract review arrangements between the laboratory and the customer	Volume	S



4296
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

STAR LAB (UK) Ltd
Issue No: 004 Issue date: 19 March 2012

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
VOLUME of liquids (See Note 1) For water delivered from a POVA.	Nominal volume Single Channel Pipettes 0.2 µl to 2 µl 2 µl to 10 µl 10 µl to 20 µl 20 µl to 100 µl 100 µl to 200 µl 200 µl to 500 µl 500 µl to 1 ml 1ml to 2.5 ml 2.5 ml to 5ml 5 ml to 10 ml Multi Channel Pipettes 1 µl to 10 µl 10 µl to 50 µl 50 µl to 300 µl	0.15 µl 0.15 µl 0.30 µl 0.60 µl 1.3 µl 2.3 µl 3.1 µl 6.7 µl 22 µl 75 µl 0.15 µl 0.60 µl 2.3 µl	Note 1. For water delivered from piston and/or plunger operated volumetric apparatus. Single Channel Pipettes and Multi Channel Pipettes Note 2. Gravimetric method 10 readings or 5 readings Note 3. Users requiring conformity to ISO 8655 should note that this cannot be demonstrated using fewer than 10 readings.	P and S
END				