


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p><b>UKAS</b> CALIBRATION</p> <p>4491</p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<h3>Oldbaum Services Limited</h3> <p><b>Issue No: 004    Issue date: 01 October 2010</b></p>	
	<p><b>Unit 5, Alpha Centre</b> <b>Innovation Park</b> <b>Stirling University</b> <b>Stirling</b> <b>FK9 4NF</b></p>	<p><b>Contact: Ms M Griesbaum</b> <b>Tel: +44 (0)1786 469639</b> <b>Fax:</b> <b>E-Mail: monica@oldbaumservices.co.uk</b> <b>Website: www.oldbaumservices.co.uk</b></p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
AIR VELOCITY Calibration of Anemometers	4 m/s 5 m/s 6 m/s 7 m/s 8 m/s 9 m/s 10 m/s 11 m/s 12 m/s 13 m/s 14 m/s 15 m/s 16 m/s 17 m/s to 18 m/s 19 m/s to 20 m/s 21 m/s to 22 m/s 23 m/s to 28 m/s 29 m/s to 35 m/s	0.88 % 0.74 % 0.68 % 0.63 % 0.61 % 0.59 % 0.57 % 0.56 % 0.66 % 0.62 % 0.60 % 0.59 % 0.58 % 0.56 % 0.55 % 0.54 % 0.53 % 0.52 %	Large Anemometers can be calibrated.
END			