


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <b>UKAS</b> CALIBRATION <b>4518</b>  Accredited to <b>ISO/IEC 17025:2005</b>	<b>M J Wilson Group Ltd</b>	
	Issue No: 003    Issue date: 04 January 2012	
	<b>Flotech House</b> Stuart Road Bredbury Stockport SK6 2SR	<b>Contact: Chris Worrall</b> Tel: +44 (0)161 406 2200 Fax: +44 (0)161 406 9196 E-Mail: enquiries@flotech.co.uk Website: www.flotech.co.uk
<b>Calibration performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k = 2$ )	Remarks
FLOW Gas flow rate	0.015 l/min to 0.5 l/min 0.5 l/min to 50 l/min	0.45 % 0.31 %	Any suitable non-corrosive gas  Calibration of flow devices with an electrical output may be undertaken
END			