


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>4533</p> <p>Accredited to ISO/IEC 17025:2005</p>	<p>Geotechnical Instruments (UK) Ltd</p> <p>Issue No: 002 Issue date: 04 August 2010</p>	
	<p>Sovereign House Queensway Leamington Spa Warwickshire CV31 3JR</p>	<p>Contact: Simon Sheridan Tel: +44 (0)1926 338111 Fax: +44 (0)1926 338110 E-Mail: service@geotech.co.uk Website: www.geotech.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
GAS ANALYSIS			
Gas analysers	Amount fraction mol/mol	Amount fraction mol/mol	
carbon dioxide	50 %	1.1 %	
	15 %	0.70 %	
	5 %	0.43 %	
methane	50 %	0.94 %	
	15 %	0.64 %	
	5 %	0.41 %	
oxygen	21 %	0.31 %	
END			