

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p>UKAS CALIBRATION</p> <p>4635</p> <p>Accredited to ISO/IEC 17025:2005</p>	<h3>Triple Red Limited</h3> <p>Issue No: 001 Issue date: 03 February 2011</p>	
	<p>Unit D4 Drakes Drive Long Crendon Industrial Estate Long Crendon Buckinhhamsire HP18 9BA</p>	<p>Contact: Ms L Halson Tel: +44 (0)1844 203651 Fax: +44 (0)1844 203650 E-Mail: halsonL@triplered.com Website: www.triplered.com</p>
<p>Calibration performed by the Organisations at the locations specified below</p>		

Locations covered by the organisation and their relevant activities

Site activities performed away from the locations listed above:

Location details	Activity	Location code
<p>The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.</p> <p>Local contact Leanne Halson</p> <p>Contact details as above</p>	<p>Time Pressure Temperature</p>	<p>S</p>



4635
Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

Triple Red Limited
Issue No: 001 Issue date: 03 February 2011

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ($k=2$)	Remarks	Location Code
PRESSURE Gas pressure (gauge)	- 0.8 bar (g) to 3 bar (g)	20 mbar		
TEMPERATURE Temperature controlled autoclaves, media preparators, incubators, sterilizers, ovens, environmental chambers, and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	30 °C to 150 °C	0.50 °C	Single and multipoint monitoring probes. Time dependent temperature profiling.	
Time Interval	10 s to 1 hour	2.0 s		
END				